

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Open Frame Switch Mode Power Supply

Name and address of the applicant

EGSTON System Electronics Eggenburg GmbH
Grafenberger Strasse 37, A-3730 Eggenburg, Austria

Name and address of the manufacturer

EGSTON System Electronics Eggenburg GmbH
Grafenberger Strasse 37, A-3730 Eggenburg, Austria

Name and address of the factory

Egston System Electronic spol.s.r.o.
Průmyslova 20, Znojmo, CZ-67015, Czech Republic*Note: When more than one factory, please report on page 2* Additional Information on page 2

Ratings and principal characteristics

See page 2.

Trademark / Brand (if any)

EGSTON

Customer's Testing Facility (CTF) Stage used

/

Model / Type Ref.

E2xFyW3_z 60
See page 2.

Additional information (if necessary may also be reported on page 2)

The risk management requirements of the standard were not addressed.

A sample of the product was tested and found to be in conformity with

 Additional Information on page 2

IEC 60601-1:2005 + A1:2012

As shown in the Test Report Ref. No. which forms part of this Certificate

T223-0001/21 (2021-01-20)


This CB Test Certificate is issued by the National Certification Body

SIQ Ljubljana, Mašera-Spasićeva ulica 10, SI-1000 Ljubljana, Slovenia
T +386 1 4778 100, F +386 1 4778 444, info@siq.si, www.siq.si

SIQ Ljubljana is accredited by Slovenian Accreditation with accreditation number CP-001 in the field of certification of products, processes and services.

Date: 2021-01-20

Signature: Bojan Pečavar



Ratings and principal characteristics:

Input: 100-240 V~; 1,4-0,7 A; 50-60 Hz

Output:

Output voltage: 12 Vdc to 24 Vdc

Output current: 2,5 A to 5,0 A

Maximum output power: 60 W

Model / Type Ref.:

E2xFyW3_z 60

Where "x" can be:

- x = Open Frame version (There is no difference between x and O)
- O = Open Frame version (There is no difference between x and O)

Where "y" can be (represent version of power supply – there is no difference between them):

- x
- y
- S
- M
- H

Where "_z" can be:

- _z or Blank (There is no difference between _z and Blank)

Additional information (if necessary)

This CB test certificate substitutes previously issued CB test certificate No. SI-6507, dated 2018-03-22, due to update of test report.

Date: 2021-01-20

Signature: Bojan Pečavar

